



ITER-India's response against the pre-bid queries clarifications sought by bidders against above mentioned tender

Tender Title: for Silicon Drift Detector System with Vacuum Accessories

1. Name of the Bidder	2. Query No.	3. Ref. Tender Part/ Section No.	4. Ref. Clause No.	5. Description of query	6. Response of ITER-India
A	1.	II. Technical specifications of Silicon Drift Detector System (Qty. 2)	F. Data Processor Unit (DPU) 22. Digital pulse processor – Built in DPU 23. Multichannel analyser-with max 8192 channels and built in DPU 24. Regulated Power supply – Built in DPU	We requested you to accept external DPU.	We cannot deviate from our requirement mentioned in the technical specification
	2.	2.4 Bidder eligibility and other applicable conditions	a. Only Class –I Local Suppliers and Class-II Local Suppliers are eligible to bid for this enquiry. Non-Local Suppliers are not eligible to bid for this tender. Government notification no. P-4501/2/2017-PP(BEII) dated 16.09.2020 shall apply for this tender.	We observed that as per tender document III Purchase Preference mentioned as Yes. Only Class-I Local Suppliers and Class-II Local Suppliers are eligible to bid for this enquiry. <u>Non-Local Suppliers are not eligible to bid for this tender.</u> Please note that for this instrument there are no local manufacturers and it is fully imported with NO LOCAL content. Any vendor for this supply will fall under “NON-LOCAL” category as per PPLC policy. Hence, we requested you to accept our participation as NON-Local suppliers.	This is a DOMESTIC tender. Hence bidders complying Class-I & Class-II criteria are only eligible for participating in the bidding process.